

Calibration Certificate N.0970_2023_RO_E

It is composed of: 8 pages

- date 07 August 2023
- addressee
- application --
- date --

Referring to:

- device Digital multimeter
- manufacturer KEYSIGHT
- model 34401A
- s.n. (asset number)
- date of measurements 07 August 2023
- due date 07 August 2024
- laboratory reference RLT_RO_2023_E

1. Traceability:

The manufacturer procedure was applied (Document PN 34401-90013, Ch.4)
Standards used to calibrate inspection, measuring and test equipment are traceable to national and international standards.
The following reference standards have been used for calibration:

Asset Number	Serial Number	Certificate	Due Date
PRIM_VRI3	MY45053553	22-0388-02	2024-02-11
PRIM_VRI2	9165015	374266	2024-02-13
PRIM_SHNT_1	16-1221	22-0388-01, 22-0388-02	2024-02-11
PRIM_TF1	06179	23-0200-02	2023-11-23

The Head of the Centre
(Eng. Marco Porpora)

